


## TOILETING SLING R121




The Toileting Sling offers superb access for quick toileting and hygiene. It is easy to place and remove in a bed or chair situation. The fully padded design supports the back, offers under thigh comfort and allows for removal of clothing. Color coded positioning straps help align the sling properly on the hanger bar.

*Note: This sling requires the patient to have FULL head and neck control. Keep arms outside of the toileting sling. Place leg sections 3-4" above the knee for best positioning.*



## STANDING SLING R130

The R130 and R131 standing slings are for residents who are partially dependent, have at least 60% weight bearing capacity, have head and neck control, are cooperative, can sit up on the edge of the bed (with or without assistance), and are able to bend at the hip, knees and ankles. Use the R130 and R131 standing slings for standing assistance, quick toileting, weight bearing practice and for transferring resident from bed to chair, or chair to bed.



## TRANSPORT SLING R134

The R134 and R136 transport slings are for transferring residents with minimal weight-bearing capacity, who have head and neck control, are cooperative, can sit up on the edge of the bed (with or without assistance), and are able to bend at the hips, knees and ankles. Use the R134 and R136 transport sling for transferring resident from bed to chair, chair to car or general in-room transport. Do not raise the resident to a full standing position while using the transport sling.

## SLING POSITIONING (RECLINED)

Place fan-folded or rolled sling under patient with grab strap away from skin. Roll patient from side to side, smoothing each half of the sling. Bed rail on opposite side should be raised. Refer to Manufacturer's Sling Owner's Manual for detailed instructions.



## SLING POSITIONING (SEATED)

For Divided Leg or Toileting Sling, slide behind seated patient until belt is at waist level or bottom of back section is at the base of the spine. Place leg sections under thighs, smoothing fabric. Attach to hanger bar in one of the three methods: crossed between legs, under both legs or single loop under each leg.

